

Mouse NGAL ELISA

Item Number: 1040
 Lot Number: 22B23K
 Expiration Date: February 2023

Specification	Requirement		Results
	Min.	Max.	
Well to Well Variation	Average CV% \leq 10%		4.6%
OD High Concentration	\leq 0.5 A		0.495
OD Low Concentration			1.886
Dose-Response	\geq 0.8A		1.391
D-R Model			Semi-log
Model Equation			$Y = -0.7742 \log X + 0.2063$
Correlation Coefficient			0.9979
Measure of Fit			0.9959
Performance	Satisfactory/Unsatisfactory		Satisfactory

Lynn Frederick

Quality Assurance

02/23/2022

Date